

**Iron  
Eisen  
Fer  
Hierro**


**114403**

**MColortest™**


**Measuring range / Messbereich / Domaine de mesure / Intervalo de medida:**

0.01 – 0.20 mg/l (de) Fe


( 0.01 – 0.02 – 0.03 – 0.04 – 0.06 – 0.08 – 0.10 – 0.15 – 0.20 mg/l (de) Fe )



20 ml of sample




5 drops of reagent  
**Fe-1**



Close the tube and mix.



Reaction time:  
3 minutes



Compare the color with  
the color scale, read off  
the result.

20 ml Probe

5 Tropfen Reagenz  
**Fe-1**

Testglas verschließen  
und mischen.

Reaktionszeit:  
3 Minuten

Farbe der Farbskala zu-  
ordnen, Messwert ables-  
sen.

20 ml d'échantillon

5 gouttes de réactif  
**Fe-1**

Boucher le tube et mé-  
langer.

Temps de réaction:  
3 minutes

Comparer la couleur  
avec l'échelle colorimé-  
trique, lire le résultat.

20 ml de muestra

5 gotas de reactivo  
**Fe-1**

Cerrar el tubo y mezclar.

Tiempo de reacción:  
3 minutos

Comparar el color con la  
escala colorimétrica, leer  
el valor de medición.